



Individual Certificate

LMS - Labor und Messgeräte GmbH Stützerbach
DIN EN ISO 9001:2015 certified

Order No.: 2018-20138
Standard: DIN ISO 3507
Operator R. Schilling
Batch: 01/18
Date: 01.03.2018

Article.No.: 61372313EC
order quantity: 30
sample: 30
reference
temperature: 20°C
your order No.: 106601

Specification: Pycnometer with thermometer,
calibrated

sample	air pressure	room temperature	liquid temperature	product No.	verified volume
18/30	716 Torr	23,3°C	23,0°C	94	10,048 ml

Labor- und Messgeräte GmbH Stützerbach
Gewerbepark "Am Wald" 2b
98693 Ilmenau - Germany